

Ürün/Primer ambalaj	Ürün kodu	Lot numarası	Son kullanım tarihi
Mueller Hinton Agar / 90 mm Ø Petri	MHA-90-P	050423-10	05/10/2023

Fiziksel kalite kontrol			
Parametre	Spesifikasiyon	Sonuç	
Görünüm	Saydam	Uygun	
Renk	Açık sarı	Uygun	
pH (25°C)	7.3 ± 0.1	Uygun	

Mikrobiyal kontaminasyon			
İnkübasyon şartları	Spesifikasiyon	Sonuç	
35 °C ± 1; 48 saat	Mikrobiyal çoğalma gözlenmemeli	Uygun	

Verimlilik				
Referans besiyeri	-			
Kontrol yöntemi	Kalitatif			
Kontrol suşu	İnkübasyon şartları	Spesifikasiyon	Karakteristik reaksiyon(lar)	Sonuç
<i>Escherichia coli</i> ATCC 25922	35 °C ± 1; 24 saat	İyi gelişim	-	Uygun
<i>Pseudomonas aeruginosa</i> ATCC 9027	35 °C ± 1; 24 saat	İyi gelişim	Yeşilimsi koloniler	Uygun
<i>Staphylococcus aureus</i> ATCC 25923	35 °C ± 1; 24 saat	İyi gelişim	-	Uygun
<i>Enterococcus faecalis</i> ATCC 51299	35 °C ± 1; 24 saat	İyi gelişim	-	Uygun

Lot kalite kontrol onayı	
Tarih	Onaylayan
10/04/2023	Semanur ERİŞ Kalite Kontrol Sorumlusu 

Referanslar

ISO 11133:2014, Microbiology of food, animal feed and water - Preparation, production, storage and performance testing of culture media

CLSI M22, Quality Control for Commercially Prepared Microbiological Culture Media, 3rd Edition

Australian Society for Microbiology, Guidelines for assuring quality of food and water microbiological culture media. Second Edition, 2014

