

Analiz Sertifikası

Sayfa 1 / 1

Ürün/Primer ambalaj	Ürün kodu	Lot numarası	Son kullanım tarihi
Plate Count Agar/90 mm Ø Petri	PCA-90-P	110423-03	11/09/2023

Fiziksel kalite kontrol		
Parametre	Spesifikasiyon	Sonuç
Görünüm	Saydam	Uygun
Renk	Sarı	Uygun
pH (25°C)	7.0 ± 0.2	Uygun

Mikrobiyal kontaminasyon		
İnkübasyon şartları	Spesifikasiyon	Sonuç
35±1°C; 48 ± 3 saat	Mikrobiyal çoğalma gözlenmemeli	Uygun

Verimlilik				
Referans besiyeri	TSA			
Kontrol yöntemi	Kantitatif			
Kontrol suşu	İnkübasyon şartları	Spesifikasiyon	Karakteristik reaksiyon(lar)	Sonuç
<i>Staphylococcus aureus</i> ATCC 25923	35±1°C; 48 ± 3 saat	P _R ≥ 0,7	-	Uygun
<i>Bacillus subtilis subsp. Spizizenii</i> ATCC 6633	35±1°C; 48 ± 3 saat	P _R ≥ 0,7	-	Uygun
<i>Escherichia coli</i> ATCC 25922	35±1°C; 48 ± 3 saat	P _R ≥ 0,7	-	Uygun

Lot kalite kontrol onayı	
Tarih	Onaylayan
13/04/2023	Semanur ERİŞ Kalite Kontrol Sorumlusu 

Referanslar

ISO 11133:2014, Microbiology of food, animal feed and water - Preparation, production, storage and performance testing of culture media

CLSI M22, Quality Control for Commercially Prepared Microbiological Culture Media, 3rd Edition

Australian Society for Microbiology, Guidelines for assuring quality of food and water microbiological culture media. Second Edition, 2014

